



Bender UK Ltd.

Low Mill Business Park, Ulverston • Cumbria LA 12 9 EE
Tel.: 01229 480 123 • Fax: 01229 480 345
info@bender-uk.com • www.bender-uk.com



UK - Declaration of Conformity

Manufacturer: **Bender GmbH & Co. KG**

declare under our sole responsibility that the product

Product name: **Insulation monitoring device isoMED527 (see annex)**

to which this declaration relates, is in conformity with the following UK legislations.

Legislations:	RoHS Regulations 2012	S.I. 2012/3032
	Radio Equipment Regulations 2017	S.I. 2017/1206

The assessment of this product has been based on the following standards:

Applied standards :

EN 61010-1	:2010 + A1:2019 + A1:2019/AC:2019
EN 61557-8	:2015/AC:2016
EN IEC 61010-2-030	:2021 + A11:2021
EN IEC 61326-2-4	:2021
EN IEC 61557-1	:2021 + A1:2024
EN IEC 63000	:2018

Gruenberg, 30.06.2026


(Alexander Doerr, Product Validation)

Remark:

The annexes are part of this UK declaration.
Limitation of standards are marked with an asterisk (*) in the attached type list.

**Bender UK Ltd.**

Low Mill Business Park, Ulverston • Cumbria LA 12 9 EE
Tel.: 01229 480 123 • Fax: 01229 480 345
info@bender-uk.com • www.bender-uk.com

**Annex - Typelist of UKCA-declaration isoMED527**

Product group	Art.-no.	Device type
isoMED527	B81625521	isoMED527-L-2
isoMED527	B81625524	isoMED527-L-2 inkl. Funktionsmodul A
isoMED527	B81625524L	isoMED527-L-2 inkl. Funktionsmodul A